

CALIBRATION REPORT



Report No. : ETPL/MFG-1/T/2022/10/75		Date: 26.10.2022	
Name of the Customer	INDO - MIM PRIVATE LIMITED 45[P] Kiadb Industrial Area Doddaballapur Bangalore , Karnataka - 561203		
Equipment Name	Precision Balance		
Model	AJ-12KE		
Capacity / Accuracy	12000g/0.1g		
Platform size	180x160mm		
Sale Order No.	600115218		
Machine Sl.No	211013094		
Min & Max Range	Min: 10g	Max: 12000 g	
Acceptance Criteria	+/- 1d (+ 0.1 or - 0.1)		
Function test	Weight placed	Display reading	Error
	3000g	3000.0	Nil
	7000g	7000.0	Nil
	12000g	12000.0	Nil
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/159/O/206/2022-23 RRSL/B/299/O/371/2022-23 RRSL/B/289/O/354/2022-23		
Next Calibration Due	25.10.2023		

Inspected by : Sathish G
Date : 26.10.2022

Approved By:

Format No: 05 049 R05