

CALIBRATION REPORT



Report No. : ETPL/MFG-1/T/2022/11/86		Date: 17.11.2022	
Name of the Customer	SAHYADRI INDUSTRIES LIMITED Plot No KK2(N) & KK2(S) Sipcot Industrial Growth Centre NH 47 Kovai Main Road P V Palayam Erode Perundurai , TamilNadu - 638052		
Equipment Name	Precision Balance		
Model	AJ-220E		
Capacity / Accuracy	220g /0.001g		
Platform size	118 mm Dia		
Sale Order No.	600116047		
Machine Sl.No	221001059		
Min & Max Range	Min: 0.1g	Max: 220g	
Acceptance Criteria	+/- 1d (+ 0.001 or - 0.001)		
Function test	Weight placed	Display reading	Error
	0.1g	0.100	Nil
	100g	100.000	Nil
	220g	220.000	Nil
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/159/O/206/2022-23		
Next Calibration Due	16.11.2023		

Inspected by : Sathish G
Date : 17.11.2022

Approved By:

Format No: 05 049 R05

Essae-Teraoka Pvt Ltd
Plot No. 4B, Bommasandra
Industrial Area, Bengaluru – 560 099