

# CALIBRATION REPORT



Report No. : ETPL/MFG-1/T/2022/11/87		Date: 17.11.2022	
Name of the Customer	<b>SAHYADRI INDUSTRIES LIMITED</b> D-III Block MIDC Chinchwad Pune , Maharashtra - 411019		
Equipment Name	Precision Balance		
Model	AJ-220E		
Capacity / Accuracy	220g /0.001g		
Platform size	118 mm Dia		
Sale Order No.	600116049		
Machine Sl.No	221001060		
Min & Max Range	Min: 0.1g	Max: 220g	
Acceptance Criteria	+/- 1d (+ 0.001 or - 0.001 )		
Function test	<b>Weight placed</b>	<b>Display reading</b>	<b>Error</b>
	0.1g	0.100	Nil
	100g	100.000	Nil
	220g	220.000	Nil
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/159/O/206/2022-23		
Next Calibration Due	16.11.2023		

Inspected by : Sathish G  
Date : 17.11.2022

Approved By:

Format No: 05 049 R05

Essae-Teraoka Pvt Ltd  
Plot No. 4B, Bommasandra  
Industrial Area, Bengaluru – 560 099