

CALIBRATION REPORT



Report No. : ETPL/MFG-1/T/2022/1/85		Date: 16.11.2022	
Name of the Customer	GRASIM INDUSTRIES LIMITED Chemical Division Karwar Binaga Karwar Uttara Kannada , Karnataka - 581307		
Equipment Name	Electronic Weighing Scale		
Model	DS-315		
Capacity / Accuracy	3t / 1kg		
Platform size	NA (Hanging)		
Sale Order No.	600116063		
Machine Sl.No	2220444		
Min & Max Range	Min: 20kg	Max: 3t	
Acceptance Criteria	+/- 1d (+ 1 or - 1)		
Function test	Weight placed	Display reading	Error
	1t	1000	Nil
	2t	2000	Nil
	3t	3000	Nil
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/738/O/889/2021-22		
Next Calibration Due	15.11.2023		

Format No: 05 049 R05

 Inspected by : Sathish G
 Date : 16.11.2022

Approved By:

 Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099