

CALIBRATION REPORT



Report No. : ETPL/MFG-1/T/2023/04/53		Date: 24.04.2023	
Name of the Customer	ARIHANT METALS AND EXTRUDED PVT LTD Unit-2 IP-25 Khed City Nimgaon Village Taluka-Khed Pune , Maharashtra - 410505		
Equipment Name	Electronic Weighing Scale		
Model	DS-315		
Capacity / Accuracy	5t / 2kg		
Platform size	NA (Hanging)		
Sale Order No.	600121689		
Machine Sl.No	2230067		
Min & Max Range	Min: 40kg	Max: 5t	
Acceptance Criteria	+/- 1d (+ 2 or - 2)		
Function test	Weight placed	Display reading	Error
	1t	1000	Nil
	3t	3000	Nil
	5t	5000	Nil
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/383/O/474/2022-23		
Next Calibration Due	23.04.2024		

Inspected by : Sathish G
Date : 24.04.2023

Approved By:

Format No: 05 049 R05

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099