

CALIBRATION REPORT



Report No. :ETPL/MFG-1/2022/11/024		Date: 09.11.2022	
Name of the Customer	GOODRICH AEROSPACE SERVICES PRIVATE LIMITED., SURVEY NO 14/1 & 15/1 MARUTHI INDUSTRIAL ESTATE PHASE 2 HOODI VILLAGE K R PURAM BANGALORE, KARNATAKA, 560048		
Equipment Name	Electronic Weighing Machine		
Model	DC-810		
Capacity / Accuracy	6 kg / 0.5 Grm		
Platform size	340 X 280 MM		
Sale Order No.	600115613		
Machine SI.No	DC81022381640		
Min & Max Range	Min: 25Grm	Max: 6kg	
Acceptance Criteria	+/- 1d (+ 0.5 or - 0.5)Grm		
Function test	Weight placed	Display reading(Grm)	Error
	25Grm	25.0	Nil
	4kg	4000.0	Nil
	6Kg	6000.5	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/299/O/371/2022-23 (26/08/2022) RRSL/B/289/O/354/2022-23(30/08/2022)		
Next Calibration Due	08.11.2023		

Format No: 05 049 R05

Inspected by : Sandeep M.C

Approved By:

Date : 09.11.2022

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099