ಕರ್ನಾಟಕ ಸರ್ಕಾರ
ಕಾಮೂನು ಮಾಪನಶಾಸ್ತ್ರ ಇಲಾಖೆ

Government of Karnataka
Department of Legal Metrology

ಮೂಲಪ್ರತಿ 10 / ನೆ 18

ಪಣ ರಸೀದಿ ನಂ/ReceiptNo 2207818262
ದಿನಾಂಕ/ Date : 17-06-2022

LCR NO : AC81121

ಸತ್ಯಾಪನ ಪ್ರಮಾಣ ಪತ್ರ / CERTIFICATE OF VERIFICATION

ಕ್ರಮ ಸಂಖ್ಯೆ / SLNo: 912022078361346
ದಿನಾಂಕ / Date: 18/06/2022

ಕರ್ನಾಟಕ ಲೀಗಲ್ ಮೆಟ್ರಾಲಜಿ (ಜಾರಿ) ನಿಯಮಗಳು 2011 ರ

ಸ್ಥಳ /ಶಿಬಿರ/ತಾಲ್ಲೂಕು
Place : (Trader Premises)

ಕಾಮೂನು ಮಾಪನಶಾಸ್ತ್ರ ಅಧಿಕಾರಿ ಹೆಸರು
Name of Legal Metrology Officer

Manojkumar

ನಿಯಮ 4 (8) (c) ಹಾಗೂ ಪೆಡ್ಯೂಲ್ VIII ರಂತೆ See Rule 4(8)(c) Schedule VIII of

The Karnataka Legal Metrology (Enforcement) Rules,2011

ಗುರುತಿನ ಸಂಖ್ಯೆ
Identification No

KNT128

ನಾನು ಈ ದಿನ ESSAE TERAOKA PVT LTD, NO-4-B, BOMMASANDRA INDUSTRIAL AREA BANGALURU ಇವರಿಗೆ ಸಂಬಂಧಪಟ್ಟ ಈ ಕೆಳಗೆ ನಮೂದಿಸಿರುವ ತೂಕ ,ಅಳತೆ ಇತ್ಯಾದಿಗಳನ್ನು ಸತ್ಯಾಪನೆ ಮಾಡಿದ್ದೇನೆ ಮತ್ತು ಮೊಹರು ಮಾಡಿದ್ದೇನೆ.

I hereby certify that i have this day verified and stamped the under mentioned weights,measures,etc., belonging to

ಪರಿಮಾಣ Quantity	ತೂಕಗಳು ,ಅಳತೆಗಳು ,ತೂಕ ಮತ್ತು ಅಳತೆ ಸಾಧನಗಳು , ಅವುಗಳ ಸಾಮರ್ಥ್ಯ ಮಾದರಿ /ಪಡೆದ , ತಯಾರಿಸಿದವರು ಹಾಗೂ ಇತರ ವಿವರಗಳು Weights,Measures,Weighing and Measuring Instruments,Capacity,Class,Manufacturer,Type etc	ಸತ್ಯಾಪನ ಶುಲ್ಕ Verification Fee Rs.	ಸಾಗಣೆ ಮತ್ತು ಇತರ ವೆಚ್ಚಗಳು Carriage,Conveyance adjusting charges etc., In Rs.
1	NAWI-E:- Capacity: 15 kg - Make: ESSAE-TERAOKA - Class: Class II - Evalue: 1g Minimum Capacity 20 e - Model No: DI-60 - Digital - Serial No: DI6022370530	500	CPP
	Grand Total ₹	500.00	

ಉಪಯೋಗದಾರರ ಸಹಿ/Used By ESSAE TERAOKA PVT LTD

ಮುಂದಿನ ಸತ್ಯಾಪನೆ ದಿನಾಂಕ /Next Verification due on **17-06-2023**
or date after repair , before being put into use, whichever is earlier.

₹ ರೂ (ಅಕ್ಷರಗಳಲ್ಲಿ) (in words) **Five Hundred only**

ಸೂಚನೆ:-
Note:-
ತಿರಸ್ಕರಿಸಿದ ಬಗ್ಗೆ ಕಾರಣಗಳೊಂದಿಗೆ ಪ್ರತ್ಯೇಕ ತಿಳುವಳಿಕೆ ಪತ್ರ ನೀಡುವುದು
1) In the case of rejected weights,measures,etc the Legal Metrology Officer shall give separate certificate of rejection mentioning the reasons of rejection against each item.
2) Please use LCR NO for further transactions.
3) As per rule 4(13) of above said rules this verification certificate should be exhibited at the trade premises in a conspicuous place.
4) This document is digitally signed, manual signature not required.

Legal MetroLogy Officer

Digitally signed By MANOJ KUMAR GAJARE, Date: 18/06/2022