



Government of Karnataka  
Department of Legal Metrology



ಕರ್ನಾಟಕ ಸರ್ಕಾರ  
ಕಾನೂನು ಮಾಪನಾಂಗ, ಇಲಾಖೆ

**ಸತ್ಯಾಪನ ಪ್ರಮಾಣಿ ಪತ್ರ**  
ಕರ್ನಾಟಕ ರೀಗಲ್ ಮೆಟ್ರಾಲಜಿ (ಪಾಂ) ನಿಯಮಗಳು 2011 ರ  
ಒಂದು 4 (B) (c) ಕಾನ್ಸೂ ಸೆಡ್ಯೂಲ್ VIII ರವೆ See Rule 4(B)(c) Schedule VIII of  
The Karnataka Legal Metrology (Enforcement) Rules, 2011

LCR NO: AC81121  
ಕ್ರಮ ಸಂಖ್ಯೆ / SLNo: 912022078345613  
ದಿನಾಂಕ / Date: 29/03/2022  
ಸ್ಥಳ / ಬಿಡುಗಡೆಗೊಳಿಸಿದ  
Place : (Trader Premises)

Place : (Trader Premises)



The Karnataka Legal Metrology (Enforcement) Rules, 2011

ವಾಸು ಈ ದಿವಿ ESSAE TERAOKA PVT LTD, NO-4-B, BOMMASANDRA INDUSTRIAL AREA BANGALURU ಇವರಿಗೆ ಸಂಬಂಧಪಟ್ಟ ಈ

ಕೆಳಗೆ ನಮೂದಿಸಿರುವ ತೂಕ, ಅಳತೆ ಇತ್ಯಾದಿಗಳನ್ನು ಸತ್ಯಾಪನೆ ಮಾಡಿದ್ದೇನೆ ಮತ್ತು ಮೊಹರು ಮಾಡಿದ್ದೇನೆ.

I hereby certify that I have this day verified and stamped the under mentioned weights, measures, etc., belonging to

ವಸ್ತುವಿನ Quantity	ತೂಕಗಳು, ಅಳತೆಗಳು, ತೂಕ ಮತ್ತು ಅಳತೆ ಸಾಧನಗಳು, Weights, Measures, Weighing and Measuring Instruments, Capacity, Class, Manufacturer, Type etc	ಸತ್ಯಾಪನ ಶುಲ್ಕ Verification Fee Rs.	ಸಾಂಸ್ಥಿಕ ವಾಕ್ಯ, ಇತರ ವಿವರಗಳು Comments/Remarks adjusting charges etc., in Rs.
1	NAWI-E.- Capacity: 300 kg - Make: Essae Teraoka - Class: Class II - Evaluate: 20g Minimum Capacity a - Model No: DS-451HP - Digital - Serial No: DS45122363335	1000	CPP
<b>Grand Total</b>		₹ 1000.00	

ಉಪरोक्तವು ಸತ್ಯವಾಗಿರುವುದನ್ನು ಸಾಬೀತುಪಡಿಸಿ ದಿನಾಂಕ /Next Verification due on 28-03-2023  
or date after repair, before being put into use, whichever is earlier.  
ಸೂಚನೆ:-  
1) In the case of rejected weights, measures, etc the Legal Metrology Officer shall give separate certificate of rejection mentioning the reasons of rejection against each item.  
2) Please use LCR NO for further transactions.  
3) As per rule 4(13) of above said rules this verification certificate should be exhibited at the trade premises in a conspicuous place.  
4) This document is digitally signed, manual signature not required.

₹ 1000 (ಒಂದು ಸಾವಿರ)

(in words) One Thousand only



Legal Metrology Officer

Digitally signed By KUMAR M S, Date: 29/03/2022