

CALIBRATION REPORT



Report No. :ETPL/MFG-1/2022/11/012		Date: 11.11.2022	
Name of the Customer		AVRA LABORATORIES PRIVATE LIMITED., PLOT NO 48 JN PHARMACITY PARAWADA VISAKHAPATNAM, ANDHRA PRADESHA, 531019	
Equipment Name		Electronic Weighing Machine	
Model		DS-451	
Capacity / Accuracy		6 kg / 0.5 Grm	
Platform size		360 X 276 MM	
Sale Order No.		600115081	
Machine SI.No		DS45122381727	
Min & Max Range		Min: 25Grm	Max: 6kg
Acceptance Criteria		+/- 1d (+ 0.5 or - 0.5)Grm	
Function test	Weight placed	Display reading(Grm)	Error
	25Grm	25.0	Nil
	4kg	4000.0	Nil
	6Kg	6000.5	+1d
Test Result		Accepted and as per standard	
Traceability Certificate No.		RRSL/B/299/O/371/2022-23 (26/08/2022) RRSL/B/289/O/354/2022-23(30/08/2022)	
Next Calibration Due		10.11.2023	

Format No: 05 049 R05

Inspected by : Sandeep M.C

Approved By:

Date : 11.11.2022

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099