

# CALIBRATION REPORT



Report No. :ETPL/MFG-1/2022/11/028	Date: 09.11.2022		
Name of the Customer	GEM PAINTS PRIVATE LIMITED., #477-B, KIABD INDUSTRIAL AREA SOMAPURA 2ND STAGE, DOBASPET, NELAMANGALA TALUK BENGALURU, KARNATAKA, 562111		
Equipment Name	Electronic Weighing Machine		
Model	FB-600		
Capacity / Accuracy	600 g / 0.005 Grm		
Platform size	110 Dia		
Sale Order No.	600115865		
Machine SI.No	FB60022381562		
Min & Max Range	Min:0.5Grm	Max:600g	
Acceptance Criteria	+/- 1d (+ 0.005 or - 0.005)Grm		
Function test	<b>Weight placed</b>	<b>Display reading(Grm)</b>	<b>Error</b>
	0.5Grm	0.500	Nil
	400g	400.000	Nil
	600g	600.005	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/299/O/371/2022-23 (26/08/2022)		
Next Calibration Due	08.11.2023		

Format No: 05 049 R05

Inspected by : Sandeep M.C

Approved By:

Date : 09.11.2022

**Essae-Teraoka Pvt Ltd**  
 Plot No. 4B, Bommasandra  
 Industrial Area, Bengaluru – 560 099