



FORM-LM-VI
[See rule 14(3) of the Goa Legal Metrology Rules, 2011]
Government of Goa
OFFICE OF THE CONTROLLER OF LEGAL METROLOGY
Certificate of Verification

Verification Certificate No. **ACLM/CL/2022/34572**

Date: **11.03.2022**

Name of the Legal Metrology Officer: **Shri A. N. Panchwadkar**

ID No.: **19**

I hereby certify that I have this day verified under mentioned weights, measures, etc. belonging to **M/s. Essae Teroka Pvt Ltd. Locality: Fatorda, Margao- Goa**, under the Act and Rules made there under.

Sr. No.	Details of weights/measures/ weighing/ measuring instruments such as Capacity /Denomination/ Class/Make/Instrument No./Model No./ K-Factor etc...	Verification fees payable (₹)	Due Fees/Situ/ Conveyance charges (₹)
1	2	3	4
	Category Name: Electronic Weighing Instrument No. of articles: 01 Make: Essae Model No.: DS-252 Accuracy Class: III Max. Capacity: 7.5/15 kg. Min. Capacity: 20 g. e-Value: 1/2 g. Sr. No.: G25222801846 Others (if any): No. of Seal: 01 Location of Seal: Display Panel Verification Officer Remarks (If any): Verified and Sealed	100/-	
	Challan/Invoice no & date: ETPL/GOA/WM-STAMP21-22, dated: 11.03.2022.		
	Total Fees (₹)	100/-	

Total Fees ₹ 100/- (In words **Rupees One Hundred Only**) deposited vide M.R. No. **053778**, dt. **11.03.2022**, Cash/Demand Draft No. 114091, dated: 11.03.2022, Bank: HDFC, Margao.

Repaired by: NEW

Next Verification due on **10.03.2023**



[Handwritten Signature]

**Signature of Legal Metrology Officer,
Assistant Controller, Legal Metrology,
Central Laboratory, Porvorim Goa**

Note:

- 1) Display the certificate on conspicuous place of the business premises.
- 2) If any weight or measure is repaired before the date of verification, the user should get it re-verified and stamped on payment of the prescribed fee, before being put into use.
- 3) In the case of rejected weights, measures etc... the Inspector shall give separate certificate of rejection mentioning the reasons of rejection against each item.

This is an Computer generated certificate