



FORM-LM-VI
[See rule 14(3) of the Goa Legal Metrology Rules, 2011]
Government of Goa
OFFICE OF THE CONTROLLER OF LEGAL METROLOGY
Certificate of Verification

Verification Certificate No. ACLM/CL/2022/36274

Date: 29.03.2022

Name of the Legal Metrology Officer: Shri A. N. Panchwadkar

ID No.: 19

I hereby certify that I have this day verified under mentioned weights, measures, etc. belonging to **M/s. Essae Teroka Pvt Ltd. Locality: Fatorda, Margao- Goa**, under the Act and Rules made there under.


| Sr. No. | Details of weights/measures/ weighing/ measuring instruments such as Capacity /Denomination/ Class/Make/Instrument No./Model No./ K-Factor etc... | Verification fees payable (₹) | Due Fees/Situ/ Conveyance charges (₹) |
|---------|--|-------------------------------|---------------------------------------|
| 1 | 2 | 3 | 4 |
| | Category Name: Electronic Weighing Instrument No. of articles: 01 Make: Essae Model No.: DS-252 Accuracy Class: III Max. Capacity: 7.5/15 kg. Min. Capacity: 20 g. e-Value: 1/2 g. Sr. No.: G25222804520 Others (if any): No. of Seal: 01 Location of Seal: Display Panel Verification Officer Remarks (If any): Verified and Sealed | 100/- | |
| | Challan/Invoice no & date: ETPL/GOA/WM-STAMP21-22, dated: 29.03.2022. | | |
| | Total Fees (₹) | 100/- | |

Total Fees ₹ 100/- (In words **Rupees One Hundred Only**) deposited vide M.R. No. 053801, dt. 29.03.2022, Cash.

Repaired by: NEW

Next Verification due on 28.03.2023




Signature of Legal Metrology Officer,
Assistant Controller, Legal Metrology,
Central Laboratory, Porvorim Goa

Note:

- 1) Display the certificate on conspicuous place of the business premises.
- 2) If any weight or measure is repaired before the date of verification, the user should get it re-verified and stamped on payment of the prescribed fee, before being put into use.
- 3) In the case of rejected weights, measures etc... the Inspector shall give separate certificate of rejection mentioning the reasons of rejection against each item.

This is an Computer generated certificate