

# CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/13		Date: 09.11.2022	
Name of the Customer	Jsb Industries Private Limited., 38km Stone Delhi Jaipur, Highway Gurugram, Haryana-122001		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	5000 Kg / 1 kg		
Platform size	1000 X 1200 MM		
Sale Order No.	600115352		
Machine Sl.No	SI85022381103		
Min & Max Range	Min: 20 kg	Max: 5000 Kg	
Acceptance Criteria	/- 1d (+ 1 or - 1 ) Kg		
Function test	<b>Weight placed</b>	<b>Display reading (Kg)</b>	<b>Error</b>
	20 kg	20	Nil
	1000 Kg	1000	Nil
	5000 Kg	5001	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/738/O/889/2021-22, (22.12.2021) Kar.Stamping, SL NO, 91202101820638, (13.12.2021)		
Next Calibration Due	08.11.2023		

Format No: 05 049 R05

 Inspected by : Santhosh Angadi  
 Date : 09.11.2022

Approved By:

 Essae Teraoka Pvt Ltd  
 Plot No. 4B, Bommasandra  
 Industrial Area, Bengaluru – 560 099