

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/32		Date: 09.11.2022	
Name of the Customer/Place	PSJ Metals., 57, Sidco Industrial Estate, Dindigul, Tamilnadu-624003		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	60 kg / 5 g		
Platform size	600 X 600 MM		
Sale Order No.	600115505		
Machine Sl.No	SI85022381470		
Min & Max Range	Min: 250 g	Max: 60 kg	
Acceptance Criteria	+/- 1d (+ 0.005 or - 0.005) kg		
Function test	Weight placed	Display reading (kg)	Error
	250 g	0.250	Nil
	20 kg	20.000	Nil
	60 kg	60.005	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/289/O/354/2022-23, (30.08.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	08.11.2023		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 09.11.2022

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099