

# CALIBRATION REPORT



Report No. :ETPL/MFG-1/2022/11/029		Date: 11.11.2022	
Name of the Customer	ZYDUS LIFESCIENCES LIMITED., PLOT NO - 203 - 213 KUNDAIM INDUSTRIAL ESTATE KUNDAIM, GOA, 403115		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	1.5 kg / 0.1 Grm		
Platform size	205 X 250 MM		
Sale Order No.	600115618		
Machine SI.No	SI85022381636		
Min & Max Range	Min: 5Grm	Max: 1.5kg	
Acceptance Criteria	+/- 1d (+ 0.1 or - 0.1)Grm		
Function test	<b>Weight placed</b>	<b>Display reading(Grm)</b>	<b>Error</b>
	5Grm	5.0	Nil
	1kg	1000.0	Nil
	1.5Kg	1500.1	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/299/O/371/2022-23 (26/08/2022)		
Next Calibration Due	10.11.2023		

Format No: 05 049 R05

Inspected by : Sandeep M.C

Approved By:

Date : 11.11.2022

**Essae-Teraoka Pvt Ltd**  
 Plot No. 4B, Bommasandra  
 Industrial Area, Bengaluru – 560 099