

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/31		Date: 09.11.2022	
Name of the Customer/Place	TATA Consumer Products Limited., Damdim Packaging Centre C/O Amalgamated Plantations Private Limited Po Damdim, Jalpaiguri, West Bengal-735209		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	50 kg / 5 g		
Platform size	450 X 550 MM		
Sale Order No.	600115591		
Machine Sl.No	SI85022382008		
Min & Max Range	Min: 100 g	Max: 50 kg	
Acceptance Criteria	+/- 1d (+ 0.005 or - 0.005) kg		
Function test	Weight placed	Display reading (kg)	Error
	100 g	0.100	Nil
	20 kg	20.000	Nil
	50 kg	50.005	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/289/O/354/2022-23, (30.08.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	08.11.2023		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 09.11.2022

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099