

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/48	Date: 11.11.2022		
Name of the Customer/Place	FDC Limited., FDC Limited, Roha.Plot No.19 & 20/2, M.I.D.C, Industrial Area, Village Dhatav, Roha, Maharashtra-402116		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	100 kg / 5 g		
Platform size	450 X 550 MM		
Sale Order No.	600115824		
Machine Sl.No	SI85022382149		
Min & Max Range	Min: 250 g	Max: 100 kg	
Acceptance Criteria	+/- 1d (+ 0.005 or - 0.005) kg		
Function test	Weight placed	Display reading (kg)	Error
	250 g	0.250	Nil
	50 kg	50.000	Nil
	100 kg	100.005	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	10.11.2023		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 11.11.2022

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099