

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/19		Date: 14.11.2022	
Name of the Customer	Sivaraj Spinning Mills Private Limited., Spinning Unit, Sf No, 201/1 Vedasandur Road, Kottaiyur, Agaram Dindigul, Tamilnadu-624709		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	600 Kg / 100 g		
Platform size	1000 X 1200 MM		
Sale Order No.	600115979		
Machine Sl.No	SI85022382477		
Min & Max Range	Min: 2 kg	Max: 600 Kg	
Acceptance Criteria	/- 1d (+ 0.1 or - 0.1) Kg		
Function test	Weight placed	Display reading (Kg)	Error
	2 kg	2.0	Nil
	200 Kg	200.0	Nil
	600 Kg	600.1	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping, SL NO, 91202101820638, (13.12.2021)		
Next Calibration Due	13.11.2023		

Format No: 05 049 R05

 Inspected by : Santhosh Angadi
 Date : 14.11.2022

Approved By:

 Essae Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099