

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/20		Date: 14.11.2022	
Name of the Customer		Sivaraj Spinning Mills Private Limited., Spinning Unit, Sf No, 201/1 Vedasandur Road, Kottaiyur, Agaram Dindigul, Tamilnadu-624709	
Equipment Name		Electronic Weighing Machine	
Model		SI-850	
Capacity / Accuracy		600 Kg / 100 g	
Platform size		1200 X 1500 MM	
Sale Order No.		600115979	
Machine Sl.No		SI85022382478	
Min & Max Range		Min: 2 kg	Max: 600 Kg
Acceptance Criteria		/- 1d (+ 0.1 or - 0.1) Kg	
Function test	Weight placed	Display reading (Kg)	Error
	2 kg	2.0	Nil
	200 Kg	200.0	Nil
	600 Kg	600.1	+1d
Test Result		Accepted and as per standard	
Traceability Certificate No.		RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping, SL NO, 91202101820638, (13.12.2021)	
Next Calibration Due		13.11.2023	

Format No: 05 049 R05

 Inspected by : Santhosh Angadi
 Date : 14.11.2022

Approved By:

 Essae Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099