

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2022/11/24		Date: 16.11.2022	
Name of the Customer	Sivaraj Spinning Mills Private Limited., Sf No, 2010/1, Vedasandur Road, Kottaiyur, Agaram Village, Dindigul Tamilnadu-624709		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	300 Kg / 50 g		
Platform size	1000 X 1200 MM		
Sale Order No.	600115995		
Machine Sl.No	SI85022382614		
Min & Max Range	Min: 1 kg	Max: 300 Kg	
Acceptance Criteria	/- 1d (+ 0.05 or - 0.05) Kg		
Function test	Weight placed	Display reading (Kg)	Error
	1 kg	1.00	Nil
	100 Kg	100.00	Nil
	300 Kg	300.05	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping, SL NO, 91202101820638, (13.12.2021)		
Next Calibration Due	15.11.2023		

Format No: 05 049 R05

 Inspected by : Santhosh Angadi
 Date : 16.11.2022

Approved By:

 Essae Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099