

# CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/44		Date: 06.04.2023	
Name of the Customer/Place	Jyothy Labs Limited., Pithampur Unit Plot No 201 Sector 1 Pithampur Industrial Area, Pithampur, Dhar, Madhya Pradesh-454775		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	150 kg / 10 g		
Platform size	550 X 650 MM		
Sale Order No.	600120992		
Machine Sl.No	SI85023396110		
Min & Max Range	Min: 500 g	Max: 150 kg	
Acceptance Criteria	+/- 1d (+ 0.01 or - 0.01 ) kg		
Function test	<b>Weight placed</b>	<b>Display reading (kg)</b>	<b>Error</b>
	500 g	0.50	Nil
	50 kg	50.00	Nil
	150 kg	150.01	+1d
Test Result	Accepted and as per standard		
Traceability Certificate No.	RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	05.04.2024		

Format No: 05 049 R05

 Inspected by : Dhamodharan M  
 Date : 06.04.2023

Approved By:

**Essae-Teraoka Pvt Ltd**  
 Plot No. 4B, Bommasandra  
 Industrial Area, Bengaluru – 560 099