

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/148		Date: 25.04.2023	
Name of the Customer/Place	India Japan Lighting Private Limited., 1 Tiruvellore High Road, Pudhuchatram(Via), Tirumazhisai (Via), Chennai, Tamilnadu-600124		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	150 kg / 10 g		
Platform size	550 X 650 MM		
Sale Order No.	600121541		
Machine Sl.No	SI85023397149		
Min & Max Range	Min: 500 g	Max: 150 kg	
Acceptance Criteria	+/- 1d (+0.01 or - 0.01) kg		
Function test	Weight placed	Display reading (kg)	Error
	500 g	0.50	Nil
	50 kg	50.00	Nil
	150 kg	150.01	+1d
Test Result	Accepted and as per Acceptance Criteria		
Traceability Certificate No.	RRSL/B/343/O/420/2022-23, (16.09.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	24.04.2024		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 25.04.2023

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099