

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/153		Date: 25.04.2023	
Name of the Customer/Place	MRF Limited., Eripakkam Nettapakkam Commune, Puducherry, Puducherry-605106		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	15 kg / 1 g		
Platform size	400 X 400 MM		
Sale Order No.	600120882		
Machine Sl.No	SI85023397206		
Min & Max Range	Min: 50 g	Max: 15 kg	
Acceptance Criteria	+/- 1d (+0.001 or - 0.001) kg		
Function test	Weight placed	Display reading (kg)	Error
	50 g	0.050	Nil
	5 kg	5.000	Nil
	15 kg	15.001	+1d
Test Result	Accepted and as per Acceptance Criteria		
Traceability Certificate No.	RRSL/B/289/O/354/2022-23, (30.08.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	24.04.2024		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 25.04.2023

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099