

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/150		Date: 25.04.2023	
Name of the Customer/Place		Caplin Point Laboratories Limited., No.85/3 Suthukeny Village Mannadipet Commune Panchayat, Pondicherry, Pondicherry-605502	
Equipment Name		Electronic Weighing Machine	
Model		SI-850	
Capacity / Accuracy		60 kg / 5 g	
Platform size		450 X 550 MM	
Sale Order No.		600121546	
Machine Sl.No		SI85023397214	
Min & Max Range		Min: 250 g	Max: 60 kg
Acceptance Criteria		+/- 1d (+0.005 or - 0.005) kg	
Function test	Weight placed	Display reading (kg)	Error
	250 g	0.250	Nil
	20 kg	20.000	Nil
	60 kg	60.005	+1d
Test Result		Accepted and as per Acceptance Criteria	
Traceability Certificate No.		RRSL/B/289/O/354/2022-23, (30.08.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)	
Next Calibration Due		24.04.2024	

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 25.04.2023

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099