

CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/152		Date: 25.04.2023	
Name of the Customer/Place	Caplin Point Laboratories Limited., No.85/3 Suthukeny Village Mannadipet Commune Panchayat, Pondicherry, Pondicherry-605502		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	60 kg / 5 g		
Platform size	450 X 550 MM		
Sale Order No.	600121546		
Machine Sl.No	SI85023397216		
Min & Max Range	Min: 250 g	Max: 60 kg	
Acceptance Criteria	+/- 1d (+0.005 or - 0.005) kg		
Function test	Weight placed	Display reading (kg)	Error
	250 g	0.250	Nil
	20 kg	20.000	Nil
	60 kg	60.005	+1d
Test Result	Accepted and as per Acceptance Criteria		
Traceability Certificate No.	RRSL/B/289/O/354/2022-23, (30.08.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	24.04.2024		

Format No: 05 049 R05

 Inspected by : Dhamodharan M
 Date : 25.04.2023

Approved By:

Essae-Teraoka Pvt Ltd
 Plot No. 4B, Bommasandra
 Industrial Area, Bengaluru – 560 099