

# CALIBRATION REPORT



Report No. : ETPL/MFG-1/2023/04/172		Date: 26.04.2023	
Name of the Customer/Place	Endurance Technologies Limited., (Suspension Dicision) Plot No 103/6 Halol-2 Halol Masvad Industrial Estate, Vadodara-Godhra Highway Gidc Holol, Panchmahal, Gujarat-389350		
Equipment Name	Electronic Weighing Machine		
Model	SI-850		
Capacity / Accuracy	150 kg / 10 g		
Platform size	550 X 650 MM		
Sale Order No.	600121179		
Machine Sl.No	SI85023397217		
Min & Max Range	Min: 500 g	Max: 150 kg	
Acceptance Criteria	+/- 1d (+0.01 or - 0.01 ) kg		
Function test	<b>Weight placed</b>	<b>Display reading (kg)</b>	<b>Error</b>
	500 g	0.50	Nil
	50 kg	50.00	Nil
	150 kg	150.01	+1d
Test Result	Accepted and as per Acceptance Criteria		
Traceability Certificate No.	RRSL/B/343/O/401/2022-23, (16.09.2022) Kar.Stamping SL No. 91202101820638, (13.12.2021)		
Next Calibration Due	25.04.2024		

Format No: 05 049 R05

 Inspected by : Dhamodharan M  
 Date : 26.04.2023

Approved By:

**Essae-Teraoka Pvt Ltd**  
 Plot No. 4B, Bommasandra  
 Industrial Area, Bengaluru – 560 099